

# Certificate

## Biocompatibility Test

Material tested:

### **Esteticor® Blancor**

Dental alloy for metal-ceramic restorations, ISO 22674 and ISO 9693

Composition w/w%:

Pd 57.29; Ag 29.00; In 11.00; Au 1.00; Sn 1.00; Ga 0.50; Ru 0.20; B 0.01

Manufacturer:

Cendres + Métaux SA  
Rue de Boujean 122  
CH-2501 Biel/Bienne

Tests:

We confirm that the following tests for determining the biocompatibility were carried out in accordance with the international standards EN ISO 10993, «Biological evaluation of medical devices» (EN ISO 10993-1, EN ISO 10993-2, EN ISO 10993-5, EN ISO 10993-10, EN ISO 10993-12) and EN ISO 7405: 2008, «Dentistry – Preclinical evaluation of biocompatibility of medical devices used in dentistry – Test methods for dental materials».

The tests were performed according to the OECD directives «Good Laboratory Practice» (GLP) by the Institute BSL Bioservice Scientific Laboratories. The tests were coordinated and monitored by Dr. Henning, Dental Engineering. The test items were produced by a commercial dental laboratory according to the manufacturing instructions of Cendres + Métaux.

### **Cytotoxicity**

The cytotoxic potential of the dental alloy was tested in vitro with L-929 fibroblasts. Method: «Test on extracts», XTT staining, EN ISO 10993-5: 1999, EN ISO 10993-12: 2007 and EN ISO 7405: 2008, (5.4.a)3).

Test result:

**Esteticor® Blancor had no cytotoxic potential.**

### **Mutagenicity**

The mutagenicity was tested with the «Reverse Mutation Assay» using bacteria *Salmonella typhimurium*, ISO 10993-3: 2003, ISO 10993-12: 2007, EN ISO 7405: 2008, (5.4.b)7).

Test result:

**Esteticor® Blancor was non-mutagenic.**

### **Allergic sensitization**

The allergic sensitization was tested in vivo: Method: Maximization Test (Magnusson-Kligman), EN ISO 10993-10: 2007, (6.3) «Tests for irritation and delayed-type hypersensitivity», EN ISO 10993-12: 2007, EN ISO 7405: 2008 (5.4.b)5).

Test result:

**Esteticor® Blancor did not cause allergic sensitization.**

Loerrach, 2009-07-03

Dr. G. Henning  
Dental Engineering  
Unterer Lichsenweg 13  
D-79541 Loerrach

