

# Certificate

## Biocompatibility Test

Material tested: **Esteticor Implant® 32**  
Dental alloy for metal-ceramic restorations, ISO 9693

Composition w/w% : Au 32.0; Pd 40.85; Ag 19.0; Sn 5.0; In 3.0; Ru 0.15 %

Manufacturer: **Cendres & Métaux SA**  
**Rue de Boujean 122**  
**CH-2501 Biel-Bienne**

Tests: We confirm that the following tests for determining the biocompatibility of the dental alloy were carried out in accordance with the international standards EN ISO 10993, «Biological evaluation of medical devices» (EN ISO 10993-1, EN ISO 10993-2, EN ISO 10993-5, EN ISO 10993-10, EN ISO 10993-12) and EN ISO 7405: 1997, «Dentistry – Preclinical evaluation of biocompatibility of medical devices used in dentistry – Test methods for dental materials». The tests were performed according to the OECD directives «Good Laboratory Practice» (GLP) by the Institute BSL Bioservice Scientific Laboratories. The tests were coordinated and monitored by Dr. Henning + Co.  
The test items were produced by a commercial dental laboratory according to the manufacturing instructions of Cendres & Métaux SA.

### Cytotoxicity

The cytotoxic potential of the dental alloy was tested in vitro with L-929 fibroblasts:

Method:

«Test on extracts», XTT staining, EN ISO 10993-5: 1999, EN ISO 10993-12: 2002 and EN ISO 7405: 1997, (5.4.a)3).

Test result: **Esteticor Implant® 32 had no cytotoxic potential.**

### Allergic sensitization

The allergic sensitization was tested in vivo: Method: Maximization Test (Magnusson-Kligman), EN ISO 10993-10: 2002, (6.3) «Tests for irritation and delayed-type hypersensitivity», EN ISO 10993-12: 2002, EN ISO 7405: 1997 (5.4.b)5), OECD 406-92 and Directive 92/69 EEC, B.6.

Test result: **Esteticor Implant® 32 did not cause allergic sensitization.**

Basel, 2005-09-03

**Dr. Henning + Co.**  
**Dental Engineering**  
**Steinenvorstadt 13**  
**CH-4051 Basel**

